



***AMENDMENT TO SPECIFICATIONS FOR  
TENDER #0271-1726  
Digital Radiography System***

***Amendment date: January 16th, 2018***

---

**Please note that the following amendments to the tender document specifications**

1. On Page 37 of the tender document which states: “The System should have the capability of performing dual energy studies using the wall stand/flat panel detector combination. “

This specification was amended to:

The system must have the ability to capture a “low –energy” image and a “high energy” image during a single examination. (Dual Energy Chest Imaging) or bone suppression technology to transform conventional chest x-ray images into an enhanced soft tissue image without bone overlay”

The following statement is being added:

**Please provide, for the three Labrador Grenfell Health sites, optional pricing for a system or systems without the Dual Energy chest imaging or bone suppression technology. The system being proposed should meet all other applicable specifications within the tender document and previous amendments.**