



**Amendment #1 to
SPECIFICATIONS FOR
TENDER # 0171-1718
SUPPLY OF EEG/EMG SYSTEM
FOR WMRH**

The specifications amended as follows:

Under the product specification (section 2.0) for an EEG system is stated 'EEG Amplifier must be IP configurable' and 'EEG amplifier must be connected via CAT5 networking cable'. This is replaced these statements with " EEG amplifier must be connected via CAT5 networking cable or USB cable".

Also the closing date is being amended to:

CLOSING DATE: 6th of September 2017

CLOSING TIME: 2:00 PM (Newfoundland Time)